



## Swell Stand

Code: [24-9262](#)

Product Group: [Expansion \(swell\) Test Equipment, Tripod Attachment](#)

For mounting Swell Dial Gauge in position on CBR mould collar.

## Standards

ASTM D1883, AASHTO T193

## Further Information

Requires dial indicator; not included, order separately

## Specification

Construction	Machined, one-piece cast aluminum.
Contact Head	Integral part of assembly; holds dial indicator.
Construction	2" I. (50.8 mm) included.
Weight	Net 1 lb. (0.4 kg).

## Accessories



### Mechanical Dial Indicator - 25 mm Range Clockwise Movement

Code: [88-4110](#)

## Spares/Consumables



## Dial Indicator - 1.0" Range

Code: 88-4100

---

## Alternatives



## Electronic Digital Indicator, 1.0"/25.4 mm range

Code: 88-4220