



Swell Stand

Code: [24-9262](#)

Product Group: [Expansion \(swell\) Test Equipment, Tripod Attachment](#)

For mounting Swell Dial Gauge in position on CBR mould collar.

Standards

ASTM D1883, AASHTO T193

Further Information

Requires dial indicator; not included, order separately

Specification

Construction	Machined, one-piece cast aluminum.
Contact Head	Integral part of assembly; holds dial indicator.
Construction	2" I. (50.8 mm) included.
Weight	Net 1 lb. (0.4 kg).

Accessories



Dial Indicator - 1.0" Range

Code: [88-4100](#)



Mechanical Dial Indicator - 25 mm Range Clockwise Movement

Code: [88-4110](#)

Alternatives



CBR Mould

Code: [24-9229](#)



Electronic Digital Indicator, 1.0"/25.4 mm range

Code: [88-4220](#)



Swell Plate

Code: [24-9260](#)