



## Heat of Hydration Apparatus ASTM C186/EN 196 220-240V 50Hz 1Ph. C of C Required to meet ASTM C186.

Code: [38-4605/01](#)

Product Group: [N/A](#)

Used to determine the heat of hydration of low heat Portland and hydraulic cement.

The apparatus consists of a Propeller and a Dewar flask contained in an insulated material and housed in a wooden box which is hinged so that the flask can be easily removed or replaced.

A second hinged wooden box contains the first one, granting a better insulation, as expressly requested by the a.m. Standards.

The Calorimeter is supplied complete with a constant speed electric stirrer, and filler glass funnel.

The standard supply does not include a thermometer.