



Autoclave

Code: [38-3800/01](#)

Product Group: [Soundness of Portland Cement by the Autoclave Method](#)

The autoclave provides high pressure steam curing of the specimens and is supplied with a safety valve pressure gauge and controlled heater unit. The pressure vessel is independently certified and certificated. A specimen rack with capacity for 12 specimens included. The unit conforms to the requirements of ASTM C151.

220-240 V 50/60 Hz 1ph

Standards

ASTM C151

Further Information

****NOT AVAILABLE FOR SALE IN THE USA****

Spares/Consumables



Lid Sealing Gasket

Code: [38-3800/13](#)