

















# IF IT'S WORTH BUILDING, IT'S WORTH TESTING

www.ele.com

This wishlist was generated on 10/08/2016, and contains the following Products:

### 38-0200/01

**ELE Flame Photometer** 

38-0320/01 Air Compressor

38-0270 Regulators for Natural Gas

#### 485-031/01

Uv/Vis Spectrophotometer with 100 Plastic Cuvettes 10X10mm Cell Holder Instructions & Mains Lead

#### 82-8500/01

Water Bath - 12 litres

#### 24-2854/01

Mechanical End-Over-End Shaker Ce with Friction Safety Device and Control Panel 220-240V 50Hz 1Ph.

80-0200/01 ELE Sieve Shaker

#### 42-1000/01

**Buoyancy Balance 6Kg X 0.1G. Supplied with Support Frame Water Tank & Suspension Hook.** 

42-1005

Wire Basket Brass with Handle Nominal 6000Cm<sup>3</sup> Capacity with 1.7mm Wire Mesh.

82-0460 Measuring Cylinder - Glass 500 ml

82-3320 Porcelain Crucible

81-0100 Spatula 100 mm

82-2660 **Funnel** 





#### 81-0220

**Aluminium Scoop Large** 

#### 520-260

Large Current Flow Meter Wading Set for Velocities 0.03 to 10 M/S.

#### 23-7462

Speedy Moisture Tester G2 Large 0-50% Moisture Range. Supplied Less Calcium Carbide

#### 24-4650

Soil Hydrometer Type 151H ASTM D422 Graduated 0.995 to 1.038 Specific Gravity.

#### 34-2800

**Bulk Density Measure 30 Litre** 

38-3005 Le Chatelier Mould

38-3200 Extensibility of Mould Apparatus BS

38-3205 Extensibility of Mould Apparatus EN

38-1200 Le Chatelier Flask


38-2015/01 Automatic Vicat Apparatus

38-2015/12 Needle Cleaning Device

38-2015/14 Printer Paper Rolls

38-2015/15 Mould Tank

38-2015/16 Consistency Plunger





#### 78-1250/01

225 Litre Drying Oven

#### 34-4570

**150mm Cube Mould 4-Part Clamp Type. Cast Iron Construction. BS EN Compliant.** 

#### 34-5230

150mm Dia Cylinder Mould 150mm Long.

#### 78-7600

**Heavy-duty Solution Balance** 

47-0011 Calibrated Spring Balance

78-5335/01 Electronic Balance - 250 g x 0.1 mg

38-2010 Vicat Frame

82-5248 Portable Thermometer 0 - +250°C

34-0300/01 Vibro Consistometer

#### 82-0380

Measuring Cylinder - Glass 100 ml

#### 34-5053

150 X 150 X 750mm Beam Mould.

#### 42-4500

75mm Aggregate Crushing Value Apparatus. Supplied Complete with Metal Measure and Tamping Rod.

#### 36-3280/01

ADR Touch 2000 BS EN Compression Machine with Digital Readout and Self Centring Platens.

#### 37-5170

EN 12390-3/4 Distance Piece - 80 mm Depth

+44 (0)1525 249 200

+1 (800) 323 1242



#### 37-5180

EN 12390-3/4 Distance Piece - 100 mm Depth

#### 37-6130

100kN Flexural (Beams) Frame. Supplied Without Specimen Bearers and Fitting Kit.

#### 37-6138

100kN Flexural Fitting Kit (ADR) Used for Connecting Flexural Frames to ADR Compression Machines.

#### 37-6131

BS 1881 Specimen Bearer Assembly 38mm Dia X 160mm Length.

#### 37-6133

Ball Seating Assembly 150mm Dia Platens.

35-1480 Standard Concrete Test Hammer

78-7140 **Triple Beam Balance** 

### 82-5420 Digital Pocket Thermometer





### **ELE Flame Photometer**

Code: 38-0200/01 Product Group: ELE Flame Photometer

Special Note: The Flame Photometer requires a source of moisture-free air at 6 litres/minute at a pressure of 1 kg/cm<sup>2</sup>. It also requires a fuel source for the flame and a regulator (see accessories).

Standards EN 196-21, ASTM C114, TEST 17

### **Further Information**

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

For 220-240 V AC, 50-60 Hz, 1 ph


#### Specification

Dimensions (I x w x h) Readout Sensitivity Na K Ca Reproducibility

**Recorder Output** 

Weight

420 x 360 x 300 mm LED three 12.5 mm digits 3 to 100 ppm 3 to 100 ppm 5 to 100 ppm (optional filter) 1% CV for 20 consecutive samples using 10 ppm, set to read 100 Nominal 1.00 V for a reading of 100.0 8 kg







### **Air Compressor**

Code: 38-0320/01 Product Group: ELE Flame Photometer

Supplies air at 6 litres/minute at a pressure of 1 kg/cm2, for use with 38-0200/01 Flame Photometer.

Standards EN 196-21, ASTM C114, TEST 17

### **Further Information**









# **Regulators for Natural Gas**


Code: 38-0270 Product Group: ELE Flame Photometer

Standards EN 196-21, ASTM C114, TEST 17

### **Further Information**







# Uv/Vis Spectrophotometer with 100 Plastic Cuvettes 10X10mm Cell Holder Instructions & Mains Lead


Code: 485-031/01 Product Group: General laboratory instrumentation

## Spares/Consumables



Spare Xenon lamp module

Code: 485-031/10









## Water Bath - 12 litres


Code: 82-8500/01 Product Group: Water Baths

# Further Information

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

# Specification

Capacity	12 litres
Working area	210 x 300 mm
Tray	Yes
Max Water Depth	130 mm
Controller	Digital
Temperature range and	0 to 99.9
stability °C	± 0.1
Electrical supply	220 – 240 V AC,
	50 - 60 Hz, 1 ph
Rated power kW	1.5
Weight	8.1 kg





### Mechanical End-Over-End Shaker Ce with Friction Safety Device and Control Panel 220-240V 50Hz 1Ph.

Code: 24-2854/01 Product Group: Gas Jar Method

Fitted with friction safety device, capable of rotating two EL24-2830 Gas Jars at approximately 50 rpm to satisfy BS 1377.

Standards BS 1377, EN 1997-2





### **ELE Sieve Shaker**

Code: 80-0200/01 Product Group: ELE Sieve Shaker

The ELE sieve shaker is powered by an electromagnetic drive that has no rotating parts to wear making it maintenance free and extremely quiet in operation.

The unit features a triple Vertical-Lateral-Rotary vibrating action that moves the sample over the sieve producing faster more efficient sieving, while the rapid vertical movements also help keep the apertures from blinding.


The shaker is ideal for laboratory or on site use. It is robust, compact and sufficiently lightweight to be portable. The separate digital microprocessor controlled console unit incorporates a keypad for setting the sieving program and is isolated from any effects of vibration from the shaker.

As standard the shaker includes, timer 0-999 minutes, adjustable vibration intensity and adjustable intermittent or continuous operation.

#### Further Information

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

The unit accepts up to ten 200 mm or 8 inch, full height, diameter sieves and lid and receiver, or up to six 300 mm or 12 inch diameter sieves and lid and receiver. Complete as specified. Dimensions  $(I \times w \times h)$  380 x 440 x 1085 mm. For 220 - 240 V AC, 50 Hz 1ph.





### Buoyancy Balance 6Kg X 0.1G. Supplied with Support Frame Water Tank & Suspension Hook.

Code: 42-1000/01

Product Group: Particle Density (specific gravity) and Water Absorption for Coarse Aggregates

The buoyancy balance system developed by ELE consists of a rigid support frame, incorporating a water tank mounted on a platform. A mechanical lifting device is used to raise the water tank through the frame height immersing the specimen suspended below the balance. The balance supplied may also be used as a standard weighing device, thus providing a versatile and comprehensive weighing system in the laboratory.6000 g x 0.1 g supplied with frame, water tank and suspension hook.

Standards EN 1097-6, EN 12697-6, BS 812, ASTM C127, AASHTO T85

### **Further Information**

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

For 220 - 240 V AC, 50 - 60 Hz, 1 ph.





# Wire Basket Brass with Handle Nominal 6000Cm<sup>3</sup> Capacity with 1.7mm Wire Mesh.

Code: 42-1005

Product Group: Particle Density (specific gravity) and Water Absorption for Coarse Aggregates

For BS 812 Relative Density. 200 mm diameter x 190 mm deep, 1.70 mm wire mesh.

## Standards

EN 1097-6, EN 12697-6, BS 812, ASTM C127, AASHTO T85

### **Further Information**







# Measuring Cylinder - Glass 500 ml

Code: 82-0460 Product Group: Measuring Cylinders, BS Soda glass, spouted, BS EN 4788.









# Porcelain Crucible

Code: 82-3320 Product Group: Bottles and Crucibles, Loss-on-Ignition

30ml capacity complete with lid.









# Spatula 100 mm

Code: 81-0100 Product Group: Spatulas



www.ele.com +44 (0) 01525 249 200





### Funnel

Code: 82-2660 Product Group: Funnels and Flasks Polythene, 200 mm diameter.









# Aluminium Scoop Large

Code: 81-0220 Product Group: Slump Test, Scoops and Shovels

### **Further Information**

\*\*Not available for sale in the USA\*\*







# Large Current Flow Meter Wading Set for Velocities 0.03 to 10 M/S.

Code: 520-260

Product Group: Flow meters

Consisting of current meter, revolution counter, wading rods, base and knob, 3 metres wading cable. All packed into carrying case.

# Further Information

Case dimensions: 850 x 180 x 220 mm



ET.F



### Speedy Moisture Tester G2 Large 0-50% Moisture Range. Supplied Less Calcium Carbide

Code: 23-7462

Product Group: Rapid method by Speedy Moisture Tester, Rapid method by Speedy Moisture Tester

Uses a technique based on the fact that water will react with calcium carbide to form a gas and that the quantity of gas formed is directly proportional to the water present. The gas pressure is indicated on a built-in pressure gauge. The pressure gauge calibrated in percentage of moisture. Has a moisture range of 0-50% with gauge divisions every 0.5%. Holds a sample size of 8 g. Includes an electronic balance of 200 g capacity x 0.1 g readability and is supplied in plastic carrying case.

Standards ASTM D4944, AASHTO T217


#### **Further Information**

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

#### Specification

Model	G2 Large
Moisture Range	0 to 50%
Gauge divisions	0.50%
Sample weight g	8
Weight complete kg	7.5

www.ele.com







## Soil Hydrometer Type 151H ASTM D422 Graduated 0.995 to 1.038 Specific Gravity.

Code: 24-4650 Product Group: ASTM Soil Hydrometers, Soil Hydrometers

Graduated 0.995 to 1.038 g/ml.

### Standards ASTM D422, ASTM E100, AASHTO T88

### Specification

Construction Scale Calibration Weight Molded, seamless. Bouyoucos; double scale. At 68°F. Net 4 oz. (113 g).





# **Bulk Density Measure 30 Litre**

Code: 34-2800 Product Group: Bulk Density Measures, Fresh Concrete - Bulk Density Measures

Manufactured from heavy gauge steel these bulk density measures comply with the requirements of either BS 812, EN 1097-3, 12350-6 or ASTM C138. All measures incorporate carrying handles as standard.

### Standards

BS 812, EN 1097-3, EN 12350-6, ASTM C138

### **Further Information**

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

For other bulk density measures see the aggregates section





### Le Chatelier Mould

Code: 38-3005

Product Group: General Expansion of Dry Hydrated Lime by the Le Chatelier Method, Soundness of Cement Paste by the Le Chatelier Method, Soundness

Comprising a split cylinder fitted with two indicator stems. The mould is supplied complete with two glass plates and a weight 100 g  $\pm$ 10 g. Three moulds required for each test.

Standards BS 6463, EN 196-3, EN 459-2

### **Further Information**





# Extensibility of Mould Apparatus BS

Code: 38-3200 Product Group: General Expansion of Dry Hydrated Lime by the Le Chatelier Method

Conforms to BS 6463. The unit comprises a metal sleeve with a hook and set screw to fit over one of the mould pointers, and a clamp to fit onto the other pointer of the mould. Supplied complete with one weight 100 g  $\pm$ 1 g.

#### Standards EN 196-3

### **Further Information**





# **Extensibility of Mould Apparatus EN**

Code: 38-3205 Product Group: General Expansion of Dry Hydrated Lime by the Le Chatelier Method

Conforms to EN196-3, 459-2. The unit comprises a metal sleeve with a hook and set screw to fit over one of the mould pointers, and a clamp to fit onto the other pointer of the mould. Supplied complete with one weight 300 g  $\pm$ 1 g.

#### Standards EN 196-3

#### **Further Information**



ET.F



### Le Chatelier Flask

Code: 38-1200 Product Group: Specific Gravity (relative density) of Hydration Cement, Le Chatelier Flask

For determining the density of hydraulic cement and lime, supplied complete with calibration certificate. Capacity 250ml. Graduated from 0 to 1ml and from 18 to 24ml graduations. Accurate to 0.05ml.


#### Standards

EN 196-6, ASTM C128, ASTM C188, E 694, AASHTO T133

### Specification

Graduations

Stopper Weight Zero mark represents 250 ml, the capacity of the large bulb; neck graduations from 0-1 ml in 0.1 ml divisions with two extra 0.1 ml graduations above the 1.0 and below the 0 marks; bulb portion of neck holds 17 ml; top portion of neck graduated from 18-24 ml in 0.1 ml divisions; 24 ml capacity. Ground glass. Net 8 oz. (227 g).





### **Automatic Vicat Apparatus**

Code: 38-2015/01 Product Group: Automatic Vicat

Complete with EN and ASTM Initial and Final needles, Consistency Plunger, Windows(R) software and RS232 cable,  $1 \times EN$  and ASTM Mould and Glass Plate. Suitable for continuous use in saturated humidity, at a controlled temperature of  $20^{\circ}C \pm 1^{\circ}C$ .

Standards ASTM C187, ASTM C191

Further Information \*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

For 220 - 240 V AC, 50 - 60 Hz, 1 ph.

### **Spares/Consumables**




**ASTM Initial and Final Set Needle** 

Code: 38-2023



Consistency Plunger Code: 38-2015/16












### **EN Initial and Final Set Needle**

Code: 38-2021









# **Needle Cleaning Device**

Code: 38-2015/12 Product Group: Automatic Vicat Accessory.

Standards ASTM C187, ASTM C191

## **Further Information**









### **Printer Paper Rolls**

Code: 38-2015/14 Product Group: Automatic Vicat Accessory: pack of 10.

Standards ASTM C187, ASTM C191

## **Further Information**







### **Mould Tank**

Code: 38-2015/15 Product Group: Automatic Vicat

Accessory: for testing samples under water as per EN 196-3, for use in temperature controlled laboratories.

Standards ASTM C187, ASTM C191

### **Further Information**









# **Consistency Plunger**

Code: 38-2015/16 Product Group: Automatic Vicat Spare.

Standards ASTM C187, ASTM C191

### **Further Information**





# 225 Litre Drying Oven

Code: 78-1250/01 Product Group: Percentage Refusal Density (PRD), Oven Drying Method, Oven Drying Method, Oven Drying Method

#### Standards

EN 12697-32, EN 13280-4, BS 598-104, BS 1377

# Specification

Capacity	225 litres
External Dimensions (I x w x h)	540 x 1040 x 940
Internal Dimensions (l x w x h)	440 x 920 x 600
Fluctuation	± 0.75°C
Rated power heater elements	2000 W
Shelves supplied	3
Shelf positions	4
Weight	80 kg







# **150mm Cube Mould 4-Part Clamp Type. Cast Iron Construction. BS EN Compliant.**

Code: 34-4570 Product Group: Cube Moulds, Cube Moulds

4-part with clamp attached base plate.

### Specification Construction

Finish Weight Bolted sides and clamped base plate Painted 42 lb (19.1 kg)






# 150mm Dia Cylinder Mould 150mm Long.

Code: 34-5230 Product Group: Cylinder Moulds

150 mm diameter x 150 mm long, complete with base plate.







# Heavy-duty Solution Balance

Code: 78-7600 Product Group: Mechanical Balances, Heavy-Duty Solution Balance

#### **Further Information**

\*\*Not available for sale in the USA\*\*

# Specification

Capacity	20,000 g
Resolution	1.0 g
Pan size (diameter)	280 mm
Overall dimensions (I x w x h)	279 x 864 x 279 mm
Weight	20.0 kg
Capacity	20 kg.
Readability	1 g.
Beams	1,000 x 100 g; 100 x 1 g; two
	sliding
	weights.
Tare	2,270 g.
Platform	Stainless steel plate; 11 " (279 mm) diam.
Damping	Magnetic.
Attchment Weights	Slotted ratio type; one 1,000 g; two 2,000 g; one 5,000 g; one
	10,000 g (included).
Overall Dimensions	34 " w. x 11 " d. x 11 " h. (864 x 279 x 279 mm).
Weight	Net 44 lbs. (19.9 kg).







# **Calibrated Spring Balance**

Code: 47-0011 Product Group: Rate of Spread of Coated Chippings

Standards EN 12272-1, BS 598-108

## **Further Information**

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*







# Electronic Balance - 250 g x 0.1 mg

Code: 78-5335/01 Product Group: Electronic Balances

## **Further Information**

\*\*Not available for sale in the USA\*\*

## Specification

Capacity resolution Pan size Mains and battery Dimensions (I x w x h) Weight 250 g x 0.1 mg 100 mm dia No 470 x 185 x 280 mm 12 kg





#### **Vicat Frame**

Code: 38-2010 Product Group: Vicat Method

Complete with consistency plunger, 10 mm diameter. Requires one initial or final set needle to make up test weight to 300g.

Standards EN 196-3, EN 13454-2, ASTM C187, ASTM C191

## Further Information


\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

## Spares/Consumables

### **Consistency Plunger**

Code: 38-2010/10









## Portable Thermometer 0 - +250°C

Code: 82-5248 Product Group: Portable Dial-type Thermometers

Complies with BS 598 for temperature measurement of bituminous mixes prior to and during laying. Graduated in 2°C divisions with a stem length of 300 mm and accurate to  $\pm$  2°C.





## **Vibro Consistometer**

Code: 34-0300/01 Product Group: Vebe Time

Comprising of vibrating table, container, slump cone, graduated rod and plate. For 220 - 240 V AC, 50 Hz, 1 ph.





# Measuring Cylinder - Glass 100 ml

Code: 82-0380 Product Group: Measuring Cylinders, BS, Volumetric Shrinkage

Soda glass, spouted, BS EN 4788.









## 150 X 150 X 750mm Beam Mould.

Code: 34-5053 Product Group: Beam Moulds Complete with base plate





## 75mm Aggregate Crushing Value Apparatus. Supplied Complete with Metal Measure and Tamping Rod.

Code: 42-4500

Product Group: Aggregate Crushing Value (ACV) and Ten Percent Fines Value (TFV)

Comprising of 75 mm nominal diameter steel cylinder, plunger and base plate supplied complete with metal measure and tamping rod.





### ADR Touch 2000 BS EN Compression Machine with Digital Readout and Self Centring Platens.

Code: 36-3280/01 Product Group: BS EN Compression Machines

The ADR Touch range of 2000 kN and 3000 kN capacity compression machines has been designed to meet the need for reliable and consistent testing. The load frame is a welded steel fabrication carrying the ball-seated upper platen. Positively located on the loading ram, which is protected from debris by a flexible cover, the lower platen is marked for the centring of cube and cylinder specimens. Self-centring lower platens for cube location are supplied as standard on EN machines and are available as an optional extra on the standard machine. The two machines for cube testing to EN standards are assembled and aligned using a special compression frame stability tester.

The dimensions of the frame allow the testing of concrete cylinders up to 320 mm long x 160 mm diameter, 150 and 100 mm square cubes, and on EN/BS machines, 200 mm square cubes. Kerbs and flagstones may also be tested on ADR machines as well as 150 mm and 100 mm square section beams to ASTM C78 using the optional 100 kN flexural frames which are connected to the power pack.

#### Standards EN 12350-6, NF P18-411

#### **Further Information**

\*\*Not available for sale in the USA\*\*

ADR Touch BS EN 2000 and 3000 kN compression machines are supplied complete with self-centring lower platen and safety gates fitted with interlock switches ready for testing 300 x 150 mm diameter cylinders.


Whilst delivering all the features and reputation of the established ADR Series with its extensive design history, the new and improved user interface provides a high quality platform for testing that enhances the performance of ELE's compression machines.

New, sophisticated electronics further the benefits of simplified operation, whilst delivering the highest levels of accuracy in testing concrete and cement/mortar samples, satisfying the needs of Quality Control Managers, Lab Managers and Technicians.

#### ACCURACY & SAVINGS

The new ADR Touch Series, with 145mm (5.7") high resolution QVGA touch screen interface and intuitive menu-driven operation, reduces the time taken to set up the machine and perform tests, reducing the time to train staff by up to 25%.

+44 (0)1525 249 200





- Up to 6 sample types can be set as favorites, enabling one-touch set up for repeat testing.
- Full, QWERTY touch pad for input of test data.

Most testing errors produce lower strength results. Noncompliant loading rates can generate errors in measured strength. The user interface includes real-time display of load vs. time, further ensuring accurate and consistent test results and providing "goodness of test" data to improve traceability in your QC operations.

#### TRACEABILITY & DATA QUALITY

The ADR range now provides improved data quality and traceability in due diligence cases - it is now possible to demonstrate traceability all the way from the machine/user to the accreditation body, increasing your reputation and peace of mind - all test results now come complete with the machine serial number attached.

• Enhanced USB data communications between PC and machine - eliminating the need for download software.

• Two gigabytes of storage memory.

• Full customization of sample sizes - stress calculations are automatically recalculated.

USER SAFETY

With full safety gates as standard, total systems diagnostics, ram run-out switches and overload warnings ensure the

BS EN 12390-4, BS EN

safety of your employees and the reliability of your machine.

#### Specification

Testing Standards

Capacity
Cubes (Concrete)
Cylinders (Concrete)
Blocks
Flexural Testing

TFV and ACV Voltage Supply Frame Class (Stability) Frame Type Max vert. Clearance Max hor. Clearance Platen Sizes 12390-3, BS EN 12390-4, BS EN 12390-5, EN 12504-1, EN 1354, EN 1521, EN 3161, EN 1338, EN 772-6, EN 13286-41, BS 3892-3, BS 187, BS 6717, BS EN ISO 7500-1, ASTM E4 2000 kN Up to 200 mm Up to 160 x 320 mm Via optional Platen Handling System Via Flexural Frame Yes 220-240V 50/60 Hz 1Ph. Tested Welded 340 mm 355 mm Lower 220mm sq, Upper 300mm dia 50 mm

# Max Ram Travel

#### Spares/Consumables









# Pressure transducer, 0-700bar, 0.05 to 10v DC Output

Code: 6014A0062









## EN 12390-3/4 Distance Piece - 80 mm Depth

Code: 37-5170 Product Group: Distance Pieces

# **Further Information**







# EN 12390-3/4 Distance Piece - 100 mm Depth

Code: 37-5180 Product Group: Distance Pieces

# Further Information






### 100kN Flexural (Beams) Frame. Supplied Without Specimen Bearers and Fitting Kit.

Code: 37-6130 Product Group: Flexural Testing for Beams

- Open sided for ease of specimen loading
- Meets EN 12390-5, 1521, 13161, 772-6 and ASTM
- C78 for standard section beams
- Optional ball seating assembly

This rigidly constructed, open sided frame is suitable for testing 100 mm and 150 mm square-section beams. The frame supports a hydraulic ram and upper sub-platen assembly incorporating a spherical seating. The upper and lower sub-platens will accept various specimen loading bearers, which are supplied separately.

Further Information \*\*Not available for sale in the USA\*\*






## 100kN Flexural Fitting Kit (ADR) Used for Connecting Flexural Frames to ADR Compression Machines.

Code: 37-6138

Product Group: Accessories to connect ELE compression machines to the flexural/transverse frames

These flexural frames can be used with a wide range of ELE compression machines. Simply select one of two fitting kits designed to link the flexural frame to either the ADR-Auto range of compression machines or any other ADR compression machine.

#### **Further Information**





## BS 1881 Specimen Bearer Assembly 38mm Dia X 160mm Length.

Code: 37-6131

Product Group: Flexural Testing for Beams

Comprising 2 self-aligning upper bearers, 1 self-aligning and 1 fixed lower bearer. Roller bearers are 38 mm diameter x 160 mm long. Suitable for 3- or 4-point flexural testing of beams.

#### **Further Information**





# Ball Seating Assembly 150mm Dia Platens.

Code: 37-6133

Product Group: Flexural Testing for Beams

Comprising a ball seating assembly with 150 mm diameter platen and a lower platen 150 mm diameter x 16 mm thick. Suitable for testing low-strength specimens.

#### **Further Information**





## **Standard Concrete Test Hammer**

Code: 35-1480 Product Group: Concrete Test Hammer, Surface Hardness

The hammer is intended for testing the quality of concrete in finished structures such as buildings and bridges. Supplied complete with carrying case and carborundum stone, the hammer is suitable for testing concrete with compressive strengths of 10 to 70 N/mm2.

### Standards ASTM C805

## Specification

Body

**Calibration Curves** 


Rubbing Stone Accuracy Carrying Case Weight Includes indicator scale, calibration curves. Rebound number vs. compressive strength. Prepares test surface. Within 15 %. Plastic. Net 3 lbs. (1.4 kg).

#### Spares/Consumables



#### **Rubbing Stone**

Code: 35-1475/10






## **Triple Beam Balance**

Code: 78-7140 Product Group: Triple Beam Scales, Vapor Emission Test Kit, Mechanical Balances

## Specification

Capacity	2610 g
Resolution	0.1 g
Pan size (diameter)	150 mm
Overall dimensions (I x w x h)	470 x 180 x 150 mm
Weight	3.2 kg
Capacity	2,610 g.
Readability	0.1 g.
Platform	Stainless steel plate; 6â <sup>III</sup> (152 mm) diam.
	except for scale with scoop models.
Damping	Magnetic.
Attchment Weights	One 500 g; two 1,000 g (included).
Dimensions	18â00 w. x 6â00 d. x 7â00 h. (45 7 x 152 x 178 mm).
Weights	Net 8 to 10 lbs. (3.6 to 4.5 kg).





## **Digital Pocket Thermometer**

Code: 82-5420 Product Group: Digital Pocket Thermometer

The stainless steel probe carries a highly accurate platinum sensor and when not in use folds back into the side of the instrument. With a clear 8 mm high LCD display this instrument is powered by a standard 12 V MN21 battery or equivalent.

### Further Information

\*\*NOT AVAILABLE FOR SALE IN THE USA\*\*

## Specification

Dimensions (I x w x h)	150 x 45 x 19 mm
Range	-49.9 to 199.9°C
Resolution	0.1°C
Accuracy	±0.4°C
Weight	105 g

