



IF IT'S WORTH BUILDING, IT'S WORTH TESTING

Contents

www.ele.com
+44 (0) 01525 249 200



This wishlist was generated on 22/02/2017, and contains the following Products:

38-1000

Blaine Air Permeability Apparatus

38-0620

Pure Mineral Oil 1 Litre Bottle.

38-0645

Reference Cement

38-0650

Filter Papers



Blaine Air Permeability Apparatus

Code: 38-1000

Product Group: Blaine Fineness Apparatus,
Determination of Fineness Blaine Apparatus

Standards

EN 196-6, EN 459-2, EN 13286-44, BS 4359-2, ASTM C204, AASHTO T153

Specification

Cell	Stainless steel; 12.70 $\pm$ 0.10 mm i.d.
Disk	0.9 $\pm$ 0.1 mm thick with 30-40 one mm diam. holes.
Plunger	Stainless steel.
Manometer	U-tube glass with ground joint, stop cock and rubber bulb assembly.
Filter Paper	12.70 mm diam.
Fluid	Pure Mineral Oil.
Weight	Net 7 lbs. (3.2 kg).

Accessories



Electronic Balance - 620 g x 0.001 g

Code: 78-5405/01



Filter Papers

Code: [38-0650](#)



Pure Mineral Oil 1 Litre Bottle.

Code: [38-0620](#)



Reference Cement

Code: [38-0645](#)

Spares/Consumables



Blaine Air Permeability Cell

Code: [38-1000/15](#)



Filter Paper

Code: [38-0652](#)



Manometer U-tube

Code: [38-1000/10](#)



Pure Mineral Oil 1 Litre Bottle.

Code: [38-0620](#)

Product Group: [Blaine Fineness Apparatus](#),
[Determination of Fineness Blaine Apparatus](#)

Standards

EN 196-21, ASTM C204



Reference Cement

Code: [38-0645](#)

Product Group: [Blaine Fineness Apparatus](#),
[Determination of Fineness Blaine Apparatus](#)

Accessory: 5g sachet.

Standards

ASTM C204, AASHTO T153



Filter Papers

Code: [38-0650](#)

Product Group: [Determination of Fineness Blaine Apparatus](#)

Accessory: 12.7mm diameter. Box of 100.

Standards

ASTM C204

Further Information

****NOT AVAILABLE FOR SALE IN THE USA****