

















IF IT'S WORTH BUILDING, IT'S WORTH TESTING

Contents



This wishlist was generated on 10/04/2017, and contains the following Products:

25-0618

Compaction Permeameter 4"

25-0619

Compaction Permeameter 6"



Compaction Permeameter 4"

Code: 25-0618

Product Group: Compaction Permeameters

Standards

ASTM D5856

Specification

Mould Machined seamless steel

tubing; plated.

Capacity 1/30 cu. ft.; 4â 🗆 i.d. x 4.584â 🗈

h. (101.6 x 11 6.4 mm).

Collar 2â^{III} h. (50.8 mm); machined

seamless steel tubing; fits either end of mold; plated.

Top Cast aluminum; with overflow

valve

assembly and water

connection.

Base Cast aluminum; with inlet/outlet

fitting.

Porous Stone 4âll diam. x 1/2âll thick (101.6

x 12.7 mm). 50-70 permeability

rating; 225 micron, average pore size.

Weights Net 12 lbs. (5.4 kg).

Accessories

www.ele.com # +44 (0)1525 249 200 # +1 (



Constant and Falling Head Accessory Kit

Code: 25-0620

Spares/Consumables



Porous Stone for 25-0618

Code: T-319

Alternatives



Compaction Permeameter 6"

Code: 25-0619

www.ele.com 🚆 +44 (0)1525 249 200 📕 +1 (800) 323 1242 Page 4 of 6



Compaction Permeameter 6"

Code: 25-0619

Product Group: Compaction Permeameters

Standards

ASTM D5856

Specification

Mould Machined seamless steel

tubing; plated.

Capacity 0.1145 cu. ft.; 6â¹⁰ i.d. x 7â¹⁰ h.

(152 .4 x 177.8 mm).

Collar 2â^{III} h. (50.8 mm); machined

seamless steel tubing; fits either end of mold; plated.

Top Cast aluminum; with overflow

valve

assembly and water

connection.

Base Cast aluminum; with inlet/outlet

fitting.

Porous Stone 5-15 /16â diam. x 1/2â thick

(15 0.8 x 12 .7 mm).

50-70 permeability rating; 225

micron,

average pore size.

Weights Net 24 lbs. (10.9 kg).

Accessories

www.ele.com +44 (0)1525 249 200

= +1 (800) 323 1242



Constant and Falling Head Accessory Kit

Code: 25-0620

Spares/Consumables



Porous Stone for 25-0619

Code: T-324

Alternatives



Compaction Permeameter 4"

Code: 25-0618

www.ele.com 🚆 +44 (0)1525 249 200 📕 +1 (800) 323 1242 Page 6 of 6